



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® CT/NG

Cepheid Catalogue Part No.: GXCT/NG-10

Kit Lot No.: 1001457664

Cartridge Lot No.: 42903

Kit Expiration Date: 2026 11 08

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94809 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA



Solna



Sunnyvale




Lodi

Functional Testing according to DI6904, Rev. AL.1

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	CT detected; NG detected	Passed
Negative	CT not detected; NG not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature



Signature of Quality Assurance, 28 NOV 24
Date

Name: Ramon De Leon

Title: Quality Systems Specialist