

Tann Tran

Quality Systems Specialist

Name:

Title:

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpe	rt® Xpress CoV-2/Flu/RSV plus
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10
Kit Lot No.: 1000581759	
Cartridge Lot No.: 06103	
Kit Expiration Date: 2023-11-12	
Legal Manufacturer	Manufacturing Facility OSolna Sunnyvale
Cepheid	Cepheid Newark Lodi IVD (B2)
904 Caribbean Drive	904 Caribbean Drive
Sunnyvale, CA 94089	Sunnyvale, CA 94089
USA	USA
Functional Testing	
Test Description	Acceptance Criteria Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE Passed
☐ If checked this documer	t is produced electronically and valid without a wet signature.
Lumshans 30NOV22	
Signature of Quality Assurance, Date	