

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

| p | rod | net | Name:        |  |
|---|-----|-----|--------------|--|
| _ |     |     | T I STATE OF |  |

**Xpert® Xpress CoV-2/Flu/RSV plus** 

Cepheid Catalogue Part No.:

XP3COV2/FLU/RSV-10

Kit Lot No.: 1000410814

Cartridge Lot No.: 07106

Kit Expiration Date: 2023-05-14

| Legal Manufacturer  | Manufacturing Facility |        | Sunnyvale      |
|---------------------|------------------------|--------|----------------|
| Cepheid             | Cepheid                | Newark | OLodi IVD (B2) |
| 904 Caribbean Drive | 904 Caribbean Drive    |        |                |
| Sunnyvale, CA 94089 | Sunnyvale, CA 94089    |        |                |

**USA** 

Functional Testing

| Test Description | Acceptance Criteria   | Test Result |
|------------------|---|-------------|
| Negative         | SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B<br>NEGATIVE;RSV NEGATIVE | Passed      |
| Positive         | SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu<br>B POSITIVE;RSV POSITIVE | Passed      |

☐ B POSITIVE;RSV POSITIVE
☐ If checked this document is produced electronically and valid without a wet signature.

**USA** 

Signature of Quality Assurance,

Date

Name: Emaela S. Quan-Tong

Title: Quality Assurance Supervisor