

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and

\mathcal{C}^{-1}	ality System Requirements, ISO 13485, European ulations (CMDR).	
Product Name: Xpert®	BCR-ABL Ultra p190	
Cepheid Catalogue Part	No.: GXBCRABLP190-CE-10	
Kit Lot No.: 100145278	7	
Cartridge Lot No.: 0310)2	
Kit Expiration Date: 202	25-12-28	
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 U Functional Testing accord	OGA	Solna Sunnyvale
Test Description	Acceptance Criteria	Test Result
Negative	NEGATIVE [Sufficient ABL transcript]	Passed
~0.02%(p190/ABL)	POSITIVE	Passed
~0.1%(p190/ABL)	POSITIVE	Passed
If checked, this docum	nent is produced electronically and therefore valid Nov 5, 2024	without a wet signature

Date

Name: Molly Doan

Title: **Quality Systems Specialist**

Signature of Quality Assurance

302-7003 Rev C_C of A BCRABLp190, CE-IVD

Final Audit Report 2024-11-05

Created: 2024-11-05

By: Molly Doan (molly.doan@cepheid.com)

Status: Signed

Transaction ID: CBJCHBCAABAA_kGU0F23MKu-slsStwyaGilgspNgnFNF

"302-7003 Rev C_C of A BCRABLp190, CE-IVD" History

Document created by Molly Doan (molly.doan@cepheid.com) 2024-11-05 - 6:00:42 PM GMT- IP address: 45.62.187.22

Document emailed to Molly Doan (molly.doan@cepheid.com) for signature 2024-11-05 - 6:01:59 PM GMT

Document e-signed by Molly Doan (molly.doan@cepheid.com)

Signature Date: 2024-11-05 - 6:02:07 PM GMT - Time Source: server- IP address: 45.62.187.22

Agreement completed. 2024-11-05 - 6:02:07 PM GMT

