

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:_ Xpert MTB/XDR		
Cepheid Catalogue Part No.: GXMTB/XDR-10		
Kit Lot No.: 1001432996		
Cartridge Lot No.: 04307		
Kit Expiration Date: 2026-03-15		
Legal Manufacturer	Manufacturing Facility	C Solna

Functional Testing

Cepheid AB

Sweden

Röntgenvägen 5

SE-17154 Solna

Test Description	Acceptance Criteria.	Test Result
Negative	MTB NOT DETECTED	Passed
Positive - Wild Type	MTB DETECTED;INH Resistance NOT DETECTED;FLQ Resistance NOT DETECTED;AMK Resistance NOT DETECTED;KAN Resistance NOT DETECTED;CAP Resistance NOT DETECTED;ETH Resistance NOT DETECTED	Passed
Positive - Mutant	MTB DETECTED;INH Resistance DETECTED;FLQ Resistance DETECTED;AMK Resistance DETECTED;KAN Resistance DETECTED;CAP Resistance DETECTED;ETH Resistance DETECTED	Passed

Cepheid

904 Caribbean Drive

Sunnyvale, CA 94089 USA

	,,,
	06/13/2024 Date
Signature of Quality Assurance	/ / Date
Name: Hang Nguyen	
Title: Quality Systems Specia	alist

Sunnyvale