

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.

Product Name: Xpert® Xpress MVP			
Cepheid Catalogue Part No.: XPRSMVF	P-10		
Kit Lot No.: 1001462593			
Cartridge Lot No.: 03201			
Kit Expiration Date: 2025 10 12			
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA	Solna  Lodi	Sunnyvale

Functional Testing according to D36990, Rev. E.1

Test Description	Acceptance Criteria	Test Result
Positive	BV POSITIVE; Candida group DETECTED; Candida glab-krus DETECTED; TV DETECTED	Passed
Negative	BV NEGATIVE; Candida group NOT DETECTED; Candida glab-krus NOT DETECTED; TV NOT DETECTED	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

21 DEC 2024

Signature of Quality Assurance

Date

Name: Nhat Dao

Title: Quality Assurance Specialist