

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:	Xpert® Carba-R	
Cepheid Catalo	ogue Part No.: GXCARBARP-CE-10	
Kit Lot No.:	1000732772	
Cartridge Lot 1	No.: 10503	
Kit Expiration	Date: 2024-06-16	

Legal Manufacturer
Cepheid

904 Caribbean Drive Sunnyvale, CA 94089 USA Manufacturing Facility

○ Solna

Sunnyvale

Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

Mstran nat	affer and
Signature of Quality Assurance	Date
Name: Ma Stacybelle Peralta	
Title: Quality Systems Specialist	