



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® TV

Cepheid Catalogue Part No.: GXTV-10

Kit Lot No.: 1001457666

Cartridge Lot No.: 22103

Kit Expiration Date: 2026 05 17

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

Solna


Sunnyvale

Lodi

Functional Testing according to D20618, Rev. M

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	TV detected	Passed
HIGH Positive	TV detected	Passed
Negative	TV not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature



Signature of Quality Assurance, 12/20/24
Date

Name: Trinh Nguyen

Title: Quality Systems Specialist