



## CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

**Product Name:** Xpert® MRSA NXG

**Cepheid Catalogue Part No.:** GXMRSA-NXG-CE-10

**Kit Lot No.:** 1001461869

**Cartridge Lot No.:** 14005

**Kit Expiration Date:** 2026-11-22

**Legal Manufacturer**

Cepheid  
904 Caribbean Drive  
Sunnyvale, CA 94089 USA

**Manufacturing Facility**

Cepheid AB  
Röntgenvägen 5  
SE-171 54 Solna  
Sweden



Solna



Sunnyvale



Lodi

***Functional Testing according to D23916, Rev. U***

<b><i>Test Description</i></b>	<b><i>Acceptance Criteria</i></b>	<b><i>Test Result</i></b>
LOW Positive	MRSA detected	Passed
HIGH Positive	MRSA detected	Passed
Negative	MRSA not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

*Samaneh Vahid Gharavi*

Samaneh Vahid Gharavi (Dec 9, 2024 13:19 GMT+1)

**Signature of Quality Assurance,**

**Date**

**Name:** Samaneh Vahid

**Title:** Senior QA Analyst