

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

	Product Name:	Xpert® Xpress Flu/RSV
--	---------------	-----------------------

Cepheid Catalogue Part No.: XPRSFLU/RSV-CN-10

Kit Lot No.: 1001454114

Cartridge Lot No.: 46607

Kit Expiration Date: 2026-09-27

<u>Legal Manufacturer</u> <u>Manufacturing Facility</u> Osolna Osunnyvale Cepheid Lodi

904 Caribbean Drive Sunnyvale, CA 94089 904 Caribbean Drive Sunnyvale, CA 94089

USA USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

[■] If checked this document is produced electronically and valid without a wet signature.

Molly Doan Oct 22, 2024

Signature of Quality Assurance, Date

Name: Molly Doan

Title: Quality Systems Specialist

301-7334 Rev F CoA XPRSFlu.RSV plus (2)

Final Audit Report 2024-10-22

Created: 2024-10-22

By: Molly Doan (molly.doan@cepheid.com)

Status: Signed

Transaction ID: CBJCHBCAABAAv79Yg43W4DQFdBFRApKYEPjE01WT4ZHT

"301-7334 Rev F CoA XPRSFlu.RSV plus (2)" History

Document created by Molly Doan (molly.doan@cepheid.com) 2024-10-22 - 8:41:27 PM GMT- IP address: 45.62.187.0

Document emailed to Molly Doan (molly.doan@cepheid.com) for signature 2024-10-22 - 8:42:06 PM GMT

Document e-signed by Molly Doan (molly.doan@cepheid.com)

Signature Date: 2024-10-22 - 8:42:14 PM GMT - Time Source: server- IP address: 45.62.187.0

Agreement completed. 2024-10-22 - 8:42:14 PM GMT

