

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name:

Xpert® Xpress CoV-2/Flu/RSV plus

Cepheid Catalogue Part No.:

XP3COV2/FLU/RSV-10

Kit Lot No.: 1001448196

Cartridge Lot No.: 61923

Kit Expiration Date: 2025 09 14

Legal Manufacturer

Manufacturing Facility

Solna Newark Sunnyvale Lodi IVD (B2)

Cepheid

904 Caribbean Drive

Sunnyvale, CA 94089

USA

Cepheid

121 N Guild Avenue

Lodi, CA 95240

USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

Thati

23 SEP 2024

Signature of Quality Assurance,

Date

Name: Nhat Dao

Title: Quality Systems Specialist