

Product Name: Xpert® Xpress CoV-2/Flu/RSV plus

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements and ISO 13485.

Instructions for Use (IFU)

Catalogue Part No.: XPRS4			t No. and Rev.: 2-8057	
Kit Lot No.: 1001459058				
Cartridge Lot No.: 63809				
Kit Expiration Date: 2025 1	1 23			
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA	Solna • Lodi	Sunnyvale	
Functional Testing according				
Test Description	Acceptance Criteria		Test Result	
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE		Passed	
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE		Passed	
	is produced electronically and valid witho	ut a wet signature.		
Signature of Quality Assurance, Date			,	
Name: Betty Thao				
Title: Quality Systems Spec	cialist			