



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® C.difficile/Epi

Cepheid Catalogue Part No.: GXCDIFF/EPI-10

Kit Lot No.: 1001462604

Cartridge Lot No.: 42302

Kit Expiration Date: 2026-06-21

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA



Solna



Sunnyvale



Lodi

Functional Testing according to D37468, Rev. E.1

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	Toxigenic C.diff POSITIVE; 027 PRESUMPTIVE POSITIVE	Passed
Negative	Toxigenic C.diff NEGATIVE; 027 PRESUMPTIVE NEGATIVE	Passed

If checked, this document is produced electronically and therefore valid without a wet signature



Signature of Quality Assurance,

12/31/24

Date

Name: Monet Garcia

Title: Quality Assurance Specialist I