

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xper	rt® Xpress Flu/RSV					
Cepheid Catalogue Part N	XPRSFLU/RSV-CE-10					
Kit Lot No.: 1001451795						
Cartridge Lot No.: 16705						
Kit Expiration Date: 2026-08-09						
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA  Functional Testing	Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden  Sunnyvale Lodi Lodi					
Test Description	Acceptance Criteria Test Result					
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE Passed					
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE Passed					
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE Passed					
■ If checked this document is produced electronically and valid without a wet signature.						

Sara	a Mu	<i>stafa 7</i> (Oct 21, 2024	bdull	а
Sara Musta	fa Abdulla	(Oct 21, 2024	13:28 GMT-	+2)
<u>.</u>	- 1	CO	104	_

Signature of Quality Assurance, Date

Name: Sara Mustafa Abdulla

Title: QA Analyst