

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Carba-R

Catalogue Part No.: GXCARBAR-CE-10

Kit Lot No.: 1001453997

Cartridge Lot No.: 12007

Kit Expiration Date: 2026-04-12

Legal Manufacturer	Manufacturing Facility	💽 Solna	O Sunnyvale
Cepheid 904 Caribbean Drive	Cepheid AB Röntgenvägen 5		
Sunnyvale, CA 94089 USA	SE-171 54 Solna Sweden		

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP1 NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

If checked this document is produced electronically and valid without a wet signature.

<u>amaneh Vahid Gharavi</u>

Signature of Quality Assurance

Date

Name: Samaneh Vahid

Title: Senior QA Analyst