

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Administration's Quality System	1 Requirements.	
Product Name: Xpert® MTB/	RIF Ultra	
Cepheid Catalogue Part No.:	GXMTB/RIF-ULTRA-50	
Kit Lot No.: 1001452738		
Cartridge Lot No.: 60704		
Kit Expiration Date: 2026-04-	26	
Legal Manufacturer Cepheid AB Röntgenvägen 5 SE-17154 Solna Sweden	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA	Solna Sunnyo Lodi
Functional Testing accordin Test Description	g to D25862, Rev. AN Acceptance Criteria	Test Result
Wild Type Control	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED or MTB DETECTED LOW; Rif Resistance NOT DETECTED or MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED or MTB DETECTED HEIGH; Rif Resistance NOT DETECTED	Passed
Mutant Control	MTB DETECTED VERY LOW;RIF Resistance DETECTED or MTB DETECTED LOW;RIF Resistance DETECTED or MTB DETECTED MEDIUM;RIF Resistance DETECTED or MTB DETECTED HIGH;RIF Resistance DETECTED	Passed
Negative	MTB NOT DETECTED	Passed