

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name:	Xpert [®] Carba-R
Cepheid Catalo	ogue Part No.: GXCARBARP-CE-10
Kit Lot No.:	1000949183
Cartridge Lot 1	No.:21501
Kit Expiration	Date: 2024-11-03

Legal Manufacturer
Cepheid

904 Caribbean Drive Sunnyvale, CA 94089 USA Manufacturing Facility Solna Sunnyvale

Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

Signature of Quality Assura	A	01Jun23 Date
Name: Molly Doan		
Title: Quality Systems	Specialist	