

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Drug Administration's dian Medical Devices I	Quality System Requirements, ISO 13485, European IVD Regulations (CMDR).	Directive and the Cana-
Product Name: Xpe	ert® BCR-ABL Ultra	
Cepheid Catalogue Pa	art No.: GXBCRABL-10	
Kit Lot No.: 100145	7465	
Cartridge Lot No.: 4	2301	
Kit Expiration Date:		
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 Functional Testing a	904 Caribbean Drive Sunnyvale, CA 94089	Solna Sunnyval
Test Description	Acceptance Criteria	Test Result
Negative	NEGATIVE [Sufficient ABL transcript]	Passed
~0.01%(IS)	POSITIVE	Passed
~0.1%(IS)	POSITIVE	Passed
Molly Doan	cument is produced electronically and therefore valid without	out a wet signature
Molly Doan (Nov 11, 2024 10:08 PST) Signature of Quality	Nov 11, 2024 Assurance Date	

Name: Molly Doan

Title: Quality Systems Specialist

301-9243 Rev B_ C of A Xpert BCR-ABL Ultra (2)

Final Audit Report 2024-11-11

Created: 2024-11-11

By: Molly Doan (molly.doan@cepheid.com)

Status: Signed

Transaction ID: CBJCHBCAABAApeJf3pMEI6PpQwqF-HJPtprOIRgut-nu

"301-9243 Rev B_ C of A Xpert BCR-ABL Ultra (2)" History

Document created by Molly Doan (molly.doan@cepheid.com) 2024-11-11 - 6:07:29 PM GMT- IP address: 45.62.187.22

Document emailed to Molly Doan (molly.doan@cepheid.com) for signature 2024-11-11 - 6:07:51 PM GMT

Document e-signed by Molly Doan (molly.doan@cepheid.com)

Signature Date: 2024-11-11 - 6:08:00 PM GMT - Time Source: server- IP address: 45.62.187.22

Agreement completed. 2024-11-11 - 6:08:00 PM GMT



