

Title:

Quality Systems Specialist

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the Japanese Pharmaceuticals and Medical Devices Law and ISO 13485.

		and ISO 13485.			
<b>Product Name:</b>	Xpert® MRSA/SA BC				
Cepheid Catalogu	e Part No.: GXMRSA-E	BC-JP-10			
Kit Lot No.: 100	1452783				
Cartridge Lot No.	: 21902				
Kit Expiration Da	ite: 2026-04-05				
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA  Functional Testing		Manufacturing Cepheld 904 Caribbean Sunnyvale, CA USA	Drive Solna		nyvale
according to D33			Final Judgement  Product Specification		
Tast Description	Aggantawag Cuitavia	Toot Posult			
Test Description	Acceptance Criteria	Test Result	l	8	
MRSA(H)POSITIVE	MRSA POSITIVE; SA POSITIVE	Passed	Sensitivity	Passed	
MRSA(H)POSITIVE MRSA(L)POSITIVE	MRSAPOSITIVE;SAPOSITIVE MRSAPOSITIVE;SAPOSITIVE	Passed Passed	Sensitivity Precision	Passed Passed	
MRSA(H)POSITIVE MRSA(L)POSITIVE MSSA(H)POSITIVE	MRSAPOSITIVE;SAPOSITIVE MRSAPOSITIVE;SAPOSITIVE MRSANEGATIVE;SAPOSITIVE	Passed Passed Passed	Sensitivity	Passed Passed	
MRSA(H)POSITIVE	MRSAPOSITIVE;SAPOSITIVE MRSAPOSITIVE;SAPOSITIVE	Passed Passed	Sensitivity Precision	Passed Passed	

## 302-0160 Rev C

Final Audit Report 2024-12-10

Created: 2024-12-10

By: Molly Doan (molly.doan@cepheid.com)

Status: Signed

Transaction ID: CBJCHBCAABAA5V\_SWdoFKQaQDfLeyXUdQgcorU0xr4hW

## "302-0160 Rev C" History

Document created by Molly Doan (molly.doan@cepheid.com)

2024-12-10 - 10:28:41 PM GMT

Document emailed to Molly Doan (molly.doan@cepheid.com) for signature 2024-12-10 - 10:29:26 PM GMT

Document e-signed by Molly Doan (molly.doan@cepheid.com)
Signature Date: 2024-12-10 - 10:29:34 PM GMT - Time Source: server

Agreement completed. 2024-12-10 - 10:29:34 PM GMT