

Low Positive

High Positive

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xper	t® Xpress Flu/RSV			
Cepheid Catalogue Part N	XPRSFLU/RSV-CE-10			
Kit Lot No.: 1001461295 Cartridge Lot No.: 17202	2			
Kit Expiration Date: 2026	6-11-29			
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden	Solna	Sun	inyvale li
Functional Testing				
Test Description	Acceptance Crit	teria		Test Result
Negative	Flu A NEGATIVE:Flu B NEGAT	Flu A NEGATIVE:Flu B NEGATIVE:RSV NEGATIVE		Passed

■ If checked this document is	produced electronically	v and valid without	a wet signature.
= 11 checked this document is	produced electronican	y and vand without	a wor bigilalare.

Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE

Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE

	maneh Vahid Gharavi h Vahid Gharavi (Dec 16, 2024 11:33 GMT+1)		
	ure of Quality Assurance,	Date	
Name:	Samaneh Vahid		
Title:	Senior QA Analyst		

Passed

Passed