

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Apert Carba-K			
Cepheid Catalogue Part No.: GXCARBAR-CE-10			
Kit Lot No.: 1001235930			
Cartridge Lot No.: 21813			
Kit Expiration Date: 2025-03-16			
Legal Manufacturer	Manufacturing Facility	Solna	Sunnyvale
Cepheid 904 Caribbean Drive	Cepheid 904 Caribbean Drive		

Functional Testing

Sunnyvale, CA 94089 USA

Test Description	Acceptance Criteria	Test Result
LOW Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
HIGH Positive	IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed
Negative	IMP1 NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed

Sunnyvale, CA 94089 USA

Tanja Reed Tanja Reed (Sep 29, 2023 09:18 PDT)	Sep 29, 2023
Signature of Quality Assurance	Date
Name: Tanja Reed	
Title: Quality Systems Specialist	