

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

| National Medical Products | Administration's Quality System Requirements. | | |
|--|--|---------------|--|
| Product Name: Xper | Xpert® Xpress Flu/RSV | | |
| Cepheid Catalogue Part N | No.: XPRSFLU/RSV-CE-10 | | |
| Kit Lot No.: 1001458939 Cartridge Lot No.: 17201 Kit Expiration Date: 202 | 1 | | |
| Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing | Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden Sur | nnyvale di | |
| Test Description | Acceptance Criteria | Test Result | |
| Negative | Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE | Passed | |
| Low Positive | Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | Passed | |
| High Positive | Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | Passed | |
| ■ If checked this document Samaneh Vahid GharaV Samaneh Vahid GharaVi (Nov 25, 2024 14:36 GMT+1) Signature of Quality Assu Name: Samaneh Vahid | | re. | |
| Title: Senior QA Analys | t | | |