

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Canadian Medical Device Regulations (CMDR).			
Product Name: Xpert® Breast Cancer STRAT4			
Cepheid Catalogue Part No.: GXBCSTRAT4-CE-10			

Kit Lot No.: 1001462370

Cartridge Lot No.: 22201

**Kit Expiration Date:** 2026-08-09

Legal ManufacturerManufacturing FacilityCepheidCepheid904 Caribbean Drive904 Caribbean DriveSunnyvale, CA 94089 USASunnyvale, CA 94089<br/>USA

Lodi

Functional Testing according to D23611 Rev. J.2

Test Description	Acceptance Criteria	Test Result
IVTRNA NEGATIVE	ESR1 NEGATIVE;PGR NEGATIVE;ERBB2 NEGATIVE;MKi67 NEGATIVE	Passed
IVTRNA LOWPOSITIVE	ESR1 POSITIVE;PGR POSITIVE;ERBB2 POSITIVE;MKi67 POSITIVE ESR1 POSITIVE;PGR POSITIVE;ERBB2 NEGATIVE;MKi67 POSITIVE	Passed
IVTRNA HIGH POSITIVE	ESR1 POSITIVE;PGR POSITIVE;ERBB2 POSITIVE;MKi67 POSITIVE	Passed
FFPE cell line	ESR1 NEGATIVE;PGR POSITIVE;ERBB2 POSITIVE;MKi67 NEGATIVE ESR1 POSITIVE;PGR POSITIVE;ERBB2 POSITIVE;MKi67 NEGATIVE ESR1 NEGATIVE;PGR POSITIVE;ERBB2 POSITIVE;MKi67 POSITIVE ESR1 POSITIVE;PGR POSITIVE;ERBB2 POSITIVE;MKi67 POSITIVE	Passed

■ If checked, this document is produced electronically and therefore valid without a wet signature

Zenny Santos

Dec 6, 2024

Signature of Quality Assurance

Date

Zenny Santos

Name:

Quality Systems Specialist

Title: