

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with

| the US Food and Drug Adr IVD Directive and the Can | ninistration's Quality System Requadian Devices Regulations. | irements, ISO 13485, | European |
|---|---|--------------------------|--------------|
| Product Name: Xpe | rt® Xpress CoV-2/Flu/RSV plus | | |
| Cepheid Catalogue Part | No.: XP3COV2/FLU/RSV- | 10 | |
| Kit Lot No.: 1000459085 | | | |
| Cartridge Lot No.: 08715 | | | |
| Kit Expiration Date: 2023 | 3-07-09 | | |
| Legal Manufacturer | Manufacturing Facility | | innyvale |
| Cepheid | Cepheid | Newark OLo | odi IVD (B2) |
| 904 Caribbean Drive | 904 Caribbean Drive | | |
| Sunnyvale, CA 94089 | Sunnyvale, CA 94089 | | |
| USA | USA | | |
| Functional Testing | | | |
| Test Description | Acceptance Criteria | | Test Result |
| Negative | SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE | | Passed |
| Positive | SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | | Passed |
| If checked this documen | t is produced electronically and val | lid without a wet signat | ure. |

| Test Description | Acceptance Criteria | Test Result |
|------------------|---|-------------|
| Negative | SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE | Passed |
| Positive | SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE | Passed |

Signature of Quality Assurance,

Name: Renee Argishti

Title: Sr. Quality Systems Specialist