



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA NXG

Cepheid Catalogue Part No.: GXMRSA-NXG-CE-10

Kit Lot No.: 1001458868

Cartridge Lot No.: 14004

Kit Expiration Date: 2026-11-15

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

● Solna Sunnyvale
Lodi

Functional Testing according to D23916, Rev. U.1

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	MRSA detected	Passed
HIGH Positive	MRSA detected	Passed
Negative	MRSA not detected	Passed

✓ If checked, this document is produced electronically and therefore valid without a wet signature

Alexander Avramidis

Alexander Avramidis (Dec 6, 2024 07:02 GMT+1)

Signature of Quality Assurance,

Date

Name: Alexander Avramidis

Title: QA Analyst