

Title: QA Analyst

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® Xpress Flu/RSV		
Cepheid Catalogue Part N	No.: XPRSFLU/RSV-CE-10	
Kit Lot No.: 1001454043	}	
Cartridge Lot No.: 17106		
Kit Expiration Date: 2026-10-18		
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing	Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden Solna Sulla	nnyvale di
Test Description	Acceptance Criteria	Test Result
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
■ If checked this document is produced electronically and valid without a wet signature.		
Sara Mustafa Abdulla (Nov 5, 2024 07:45 GMT+1)		
Signature of Quality Assurance, Date		
Name: Sara Mustafa Abdulla		