

Signature of Quality Assurance,

Title: Quality Systems Specialist

Name: Rhoziel Cusi

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

IVD Directive and the Canad	ian Devices Regulations.		Вигореин				
Product Name: Xpert® Xpress CoV-2/Flu/RSV plus							
Cepheid Catalogue Part No	XP3COV2/FLU/RSV-	10					
Kit Lot No.: 1000604225							
Cartridge Lot No.: 06302							
Kit Expiration Date: 2023 1	1 26						
Legal Manufacturer	Manufacturing Facility		innyvale				
Cepheid	Cepheid	Newark OLo	odi IVD (B2)				
904 Caribbean Drive	904 Caribbean Drive						
Sunnyvale, CA 94089	Sunnyvale, CA 94089						
USA	USA						
Functional Testing							
Test Description	Acceptance Criteria		Test Result				
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE		Passed				
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE		Passed				
If checked this document is	s produced electronically and va	lid without a wet signar	ture.				

	0020020020	TERROR ON IVE	142,000	100
DNI	302	-7224	Day	-
PIN	3117.	-///4	K PV	•