



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR) and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® SA Nasal Complete

Cepheid Catalogue Part No.: GXSACOMP-CN-10

Kit Lot No.: 1001452838

Cartridge Lot No.: 16304

Kit Expiration Date: 2026-01-25

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Functional Testing according to D36872, Rev. H

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive Control	MRSA POSITIVE;SA POSITIVE	Passed
Negative Control	MRSA NEGATIVE;SA NEGATIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

 2024-10-23
Signature of Quality Assurance, Date

Name: Alexander Avramidis

Title: QA Analyst