

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® Carba-R					
Cepheid Catalogue Part No.: GXCARB	AR-10 •				
Kit Lot No.: 1001447214					
Cartridge Lot No.: 23601					
Kit Expiration Date: 2026-03-08					
Legal Manufacturer	Manufacturing Facility				
Cepheid	Cepheid	0	Solna	0	Sunnyval
904 Caribbean Drive	121 N Guild Avenue		Soma		Sumyva
Sunnyvale, CA 94089 USA	Lodi, CA 95240				
	USA	•	Lodi		

Functional Testing according to D18272, Rev. AF

Test Description	Acceptance Criteria	Test Result	
Low Positive	IMP-DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed	
High Positive	IMP DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED	Passed	
Negative	IMP NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED	Passed	

If checked, this document is produced electronically and therefore valid without a wet signature

Signature of Quality Assurance,

Date /

Name: Kimberly Perez

Title: Quality Systems Specialist