



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® BCR-ABL Ultra

Cepheid Catalogue Part No.: GXBCRABL-10

Kit Lot No.: 1001457622

Cartridge Lot No.: 42701

Kit Expiration Date: 2025-12-28

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089
USA



Solna



Sunnyvale



Lodi

Functional Testing according to D16624, Rev. R.1

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Negative	NEGATIVE [Sufficient ABL transcript]	Passed
~0.01%(IS)	POSITIVE	Passed
~0.1%(IS)	POSITIVE	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

ThuyTien Nguyen
ThuyTien Nguyen (Nov 30, 2024 14:58 PST)
Signature of Quality Assurance

Nov 30, 2024
Date

Name: ThuyTien Nguyen

Title: Quality Systems Specialist





GXBCRABL-10 1001457622

Final Audit Report

2024-11-30

Created:	2024-11-30
By:	ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
Status:	Signed
Transaction ID:	CBJCHBCAABAAuU9qSEbRczrTey3-EUaDn8eOaakH1W91

"GXBCRABL-10 1001457622" History

-  Document created by ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
2024-11-30 - 10:57:30 PM GMT - IP address: 216.194.108.124
-  Document emailed to ThuyTien Nguyen (thuytien.nguyen@cepheid.com) for signature
2024-11-30 - 10:58:17 PM GMT
-  Document e-signed by ThuyTien Nguyen (thuytien.nguyen@cepheid.com)
Signature Date: 2024-11-30 - 10:58:31 PM GMT - Time Source: server- IP address: 216.194.108.124
-  Agreement completed.
2024-11-30 - 10:58:31 PM GMT