

Title: Quality System Specialist

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpo	ert® Xpress CoV-2/Flu/RSV plus	
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10	
Kit Lot No.: 1000581752		
Cartridge Lot No.: 06301		
Kit Expiration Date: 2023-11-12		
Legal Manufacturer		nnyvale
Cepheid	Cepheid Newark Lo	di IVD (B2)
904 Caribbean Drive	904 Caribbean Drive	
Sunnyvale, CA 94089	Sunnyvale, CA 94089	
USA	USA	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
☐If checked this docume	nt is produced electronically and valid without a wet signatu	ure.
Signature of Quality Assurance, Date		
Name: Austin Mei		