

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Device Regulations (CMDR).

Product Name: Xpert MTB/XDR			
Cepheid Catalogue Part No.: GXMTB/XDR-10			
Kit Lot No.: 1001446019			
Cartridge Lot No.: 10301			
Kit Expiration Date: 2026-03-15			
Legal Manufacturer	Manufacturing Facility	Solna	(

Cepheid AB

Röntgenvägen 5 SE-17154 Solna Sweden

Sunnyvale

Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	MTB NOT DETECTED	Passed
Positive - Wild Type	MTB DETECTED;INH Resistance NOT DETECTED;FLQ Resistance NOT DETECTED;AMK Resistance NOT DETECTED;KAN Resistance NOT DETECTED;CAP Resistance NOT DETECTED;ETH Resistance NOT DETECTED	Passed
Positive - Mutant	MTB DETECTED;INH Resistance DETECTED;FLQ Resistance DETECTED;AMK Resistance DETECTED;KAN Resistance DETECTED;CAP Resistance DETECTED;ETH Resistance DETECTED	Passed

	DETECTE	J,EIF
III	2024-09-20	
Signature of Quality A	ssurance	Date
Name: Alexander A	Avramidis	
Title: QA Analyst		