

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Cepheid Catalogue Part No.: GXMTB/RIF-US-10

Kit Lot No.: 1001458283

Cartridge Lot No.: 44801

**Kit Expiration Date:** 2026-11-01

## Legal Manufacturer

Cepheid

904 Caribbean Drive

Sunnyvale, CA 94089 USA

## **Manufacturing Facility**

Cepheid

121 N Guild Avenue

Lodi, CA 95240

USA

) Solna

Sunnyvale

Lodi

## Functional Testing according to D31503 Rev. aa.1

Test Description	Acceptance Criteria	Test Result
Positive	MTB Detected; Rif Resistance not detected	Passed
Negative	MTB not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

11

Signature of Quality Assurance,

Date

Name: karla camarena

Title: Quality Systems Specialist