

Product Name:

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Xpert® HIV-1 Qual XC

Kit Lot No.: 1001	458982		
Cartridge Lot No.:	14501		
Kit Expiration Date:	2026-05-24		
<u>Legal Manufacturer</u> Cepheid AB		Manufacturing Facility	Solna Sunnyvale
Röntgenvägen 5 SE-17154 Solna		Cepheid AB Röntgenvägen 5	
Sweden		SE-171 54 Solna	

Sweden

Functional Testing according to D36985 Rev: C

Cepheid Catalogue Part No.: GXHIV-QA-XC-CE-10

Test Description	Acceptance Criteria	Test Result
HIV-1 Qual XC PC Low	HIV-1 DETECTED	Passed
Negative	HIV-1 NOT DETECTED	Passed

 $\hfill \square$ If checked this document is produced electronically and valid without a wet signature.

Anton Antersson 2024-12-09
Signature of Quality Assurance,

Date

Name: Anton Andersson Engström

Title: QA Analyst