

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

200	in the second	To T
Prod	uct	Name:

Xpert® NPM1 Mutation

Cepheid Catalog Part No.:

GXNPM1-CE-10

Kit Lot No .:

1001471161

Cartridge Lot No.:

03201

Kit Expiration Date:

2025-07-27

Legal	Mani	rfact	HEAR
Legal	Maii	macı	urer

Manufacturing Facility

Sunnyvale Newark

Cepheid

Cepheid

904 Caribbean Drive

904 Caribbean Drive

Sunnyvale, CA 94089 USA

Sunnyvale, CA 94089

USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Positive	NPM1 Mutation DETECTED	Passed
Negative	NPM1 Mutation NOT DETECTED [Sufficient ABL Transcript]	Passed
~.2% (NPM1/ABL)	NPM1 Mutation DETECTED	Passed
~ 1% (NPM1/ABL)	NPM1 Mutation DETECTED	Passed

Solna

Name: Molly Doan

Quality Systems Specialist

© 2022 Cepheid. All rights reserved.

PN 302-8219 Rev B