

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xper	rt® Xpress Flu/RSV	
Cepheid Catalogue Part N	No.: XPRSFLU/RSV-CE-10	
Kit Lot No.: 1000501208	8	
Cartridge Lot No.: 4241		
Kit Expiration Date: 202	24-03-31	
Legal Manufacturer	Manufacturing Facility OSolna Sunnyvale	
Cepheid	Cepheid ONewark OLodi	
904 Caribbean Drive	904 Caribbean Drive	
Sunnyvale, CA 94089	Sunnyvale, CA 94089	
USA	USA	
Functional Testing		
Test Description	Acceptance Criteria Test Result	
Negative	Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE Passed	
Low Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE Passed	
High Positive	Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE Passed	
☐ If checked this document	t is produced electronically and valid without a wet signature.	
Maria Vare		
100 pc 100ge	mi 09/11/22	
Signature of Quality Assu	rance, Date	
Signature of Quality Assu Name: Roya Kazemi	rance, Date	