

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert[®] HCV VL Fingerstick

Cepheid Catalogue Part No.: GXHCV-FS-CE-10

Kit Lot No.: 1001460186

Cartridge Lot No.: 15803

Kit Expiration Date: 2025-11-16

Legal Manufacturer

Cepheid AB
Röntgenvägen 5
SE-17154 Solna
Sweden

Manufacturing Facility

Cepheid AB
Röntgenvägen 5
SE-171 54 Solna
Sweden

Solna Sunnyvale

Functional Testing according to D51113 Rev: B

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
LOW Positive	HCV DETECTED: 2.29 - 2.67 log ₁₀ IU/mL	Passed
HIGH Positive	HCV DETECTED: 5.50 - 5.90 log ₁₀ IU/mL	Passed
Negative	HCV NOT DETECTED	Passed

If checked this document is produced electronically and valid without a wet signature.

Lava Maroof

Lava Maroof (Dec 2, 2024 07:48 GMT+1)

Signature of Quality Assurance

Date

Name: Lava Maroof

Title: QA Analyst