

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® BCR-ABL Ultra p190

Cepheid Catalogue Part No.: GXBCRABLP190-CE-10

Kit Lot No.: 1001447779

Cartridge Lot No.: 02703

Kit Expiration Date: 2025 04 06

Legal Manufacturer

904 Caribbean Drive

**Manufacturing Facility** 

Cepheid

Cepheid

Solna

Sunnyvale

Sunnyvale, CA 94089 USA

Lodi, CA 95240

121 N Guild Avenue

**USA** 

Lodi

Functional Testing according to D38418, Rev.

Test Description	Acceptance Criteria	Test Result
Negative	NEGATIVE [Sufficient ABL transcript]	Passed
~0.02% (p190/ABL)	POSITIVE	Passed
~0.1% (p190/ABL)	POSITIVE	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

Signature of Quality Assurance

Name: Jennifer Nguyen

Title:

Quality Systems Specialist