

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the

production and testing	specification Administrat	ns required by Cepheid's Cion's Quality System Requevices Regulations.	Quality System, in	compl	liance with
Product Name:	Xpert® Xpi	ess CoV-2/Flu/RSV plus			
Cepheid Catalogue l	Part No.:	XP3COV2/FLU/RSV-	10		
Kit Lot No.: 1000414 Cartridge Lot No.: 0 Kit Expiration Date:	7402				
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA Functional Testing	- - 9 S	Manufacturing Facility Sepheid 04 Caribbean Drive unnyvale, CA 94089 JSA	OSolna ONewark		nnyvale di IVD (B2)
Test Description		Acceptance Criteria			Test Result
Negative	SA	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE			Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE				Passed
☐ If checked this docu	iment is prod	luced electronically and va	lid without a wet	signati	ire.

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

Eguand 0/70122 Signature of Quality Assurance, Date Name: Emaela S. Quan-Tong Title: Quality Assurance Supervisor