

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

IVD Directive and the Cana	adian Devices Regulations.	unements, 150 15465	, European	
Product Name: Xpe	rt® Xpress CoV-2/Flu/RSV plus	<b>}</b>		
Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10				
Kit Lot No.: 1000384966				
Cartridge Lot No.: 05631				
Kit Expiration Date: 2023-01-29				
Legal Manufacturer	<b>Manufacturing Facility</b>		Sunnyvale	
Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Newark Ol	Lodi IVD (B2)	
Functional Testing				
Test Description	Acceptance Criteria		Test Result	
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE		Passed	
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE		Passed	
If checked this document is produced electronically and valid without a wet signature.				
	07/08/2022			

**Date** 

Name: Jenny Nguyen		
Title:	Quality Assurance Supervisor	

Signature of Quality Assurance,