

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpert® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10		
Kit Lot No.: 1000688898		
Cartridge Lot No.: 26201		
Kit Expiration Date: 2024 01 07		
Legal Manufacturer	The state of the s	nnyvale
Cepheid	Cepheid Newark OLo	di IVD (B2)
904 Caribbean Drive	904 Caribbean Drive	
Sunnyvale, CA 94089	Sunnyvale, CA 94089	
	USA	
USA	USA	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
☐ If checked this document is produced electronically and valid without a wet signature.		
Signature of Quality Assurance, Date		
Name: Pauline Vu		
Title: Quality Systems Specialist		