

Title: Quality System Specialist

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xp	ert® Xpress CoV-2/Flu/RSV plus	
Cepheid Catalogue Part No.: XP3COV2/FLU/RSV-10		
Kit Lot No.: 1001405352		
Cartridge Lot No.: 41014		
Kit Expiration Date: 2024-09-22		
Legal Manufacturer		nnyvale
Cepheid	Cepheid Newark Loc	li IVD (B2)
904 Caribbean Drive	904 Caribbean Drive	
Sunnyvale, CA 94089	Sunnyvale, CA 94089	
USA	USA	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
☐ If checked this document is produced electronically and valid without a wet signature.		
Roza Kazenii 01/02/24		
Signature of Quality Assurance, Date		
Name: Roya Kazemi		