

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive, the Canadian Medical Devices Regulations (CMDR), and the China National Medical Products Administration's Quality System Requirements.

Product Name: Xpert® MTB/RIF Ult	ra				
Cepheid Catalogue Part No.: GXMT	B/RIF-ULTRA-50				
Kit Lot No.: 1001457624					
Cartridge Lot No.: 61203					
Kit Expiration Date: 2026-05-10					
Legal Manufacturer	<b>Manufacturing Facility</b>				
Cepheid AB	Cepheid		Solna		Sunnyvale
Röntgenvägen 5	121 N Guild Avenue	O Some		0 0	
SE-17154 Solna	Lodi, CA 95240				
Sweden	USA	$\odot$	Lodi		

## Functional Testing according to D25862, Rev. AN

Test Description	Acceptance Criteria	Passed Passed	
Wild Type Control	MTB DETECTED VERY LOW; Rif Resistance NOT DETECTED of MTB DETECTED LOW; Rif Resistance NOT DETECTED or MTB DETECTED MEDIUM; Rif Resistance NOT DETECTED or MTB DETECTED HIGH; Rif Resistance NOT DETECTED		
Mutant Control	MTB DETECTED VERY LOW.RIF Resistance DETECTED  or MTB DETECTED LOW.RIF Resistance DETECTED  or MTB DETECTED MEDIUM.RIF Resistance DETECTED  or MTB DETECTED HIGH.RIF Resistance DETECTED		
Negative	MTB NOT DETECTED	Passed	

Mutan	nt Control	Passed			
N	legative	. M	Passed		
If ch	nesked, this do	cument is produced elect	tronically and therefore valid with $11/25/24$	out a wet signature	
Signat	ure of Quality	Assurance,	Date		
Name:	Roya Kazen	ni			
Title:	Quality Ass	urance Lead			