



CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® SA Nasal Complete

Cepheid Catalogue Part No.: GXSACOMP-10

Kit Lot No.: 1001462613

Cartridge Lot No.: 29201

Kit Expiration Date: 2026 12 13

Legal Manufacturer

Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility

Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

Solna

Sunnyvale

Lodi

Functional Testing according to D36872 Rev. H

<i>Test Description</i>	<i>Acceptance Criteria</i>	<i>Test Result</i>
Positive	MRSA Positive; SA Positive	Passed
Negative	MRSA Negative; SA Negative	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

02 JAN 2025

Signature of Quality Assurance,

Date

Name: Nhat Dao

Title: Quality Assurance Specialist I