

## **CERTIFICATE OF ANALYSIS**

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

	E	,				
Product Name: Xpe	rt® CT/NG					
Cepheid Catalogue I	Part No.: GXCT/N	IG-10				
Kit Lot No.: 10014	57792					
Cartridge Lot No.:	36210					
Kit Expiration Date:	2026-11-01					
Legal Manufacture Cepheid 904 Caribbean Drive Sunnyvale, CA 9486  Functional Testing	e 09 USA	Manufacturing Facility Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden  6904, Rev. AL.1	<ul><li>C</li></ul>	) Solna	0	Sunnyval
Test Description		Acceptance Criteria		Test Result		
Positive		CT detected; NG detected		Passed		
Negative		CT not detected; NG not detected		Passed		

■ If checked, this document is produced electronically and therefore valid without a wet signature Sara Mustafa Abdulla

Signature of Quality Assurance,	Date

Name: Sara Mustafa Abdulla

Title: QA Analyst