

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name: Xpert® Xpress CoV-2/Flu/RSV plus		
Cepheid Catalogue Part	No.: XP3COV2/FLU/RSV-10	
Kit Lot No.: 1000522483		
Cartridge Lot No.: 20614		
Kit Expiration Date: 2023-06-25		
Legal Manufacturer		innyvale
Cepheid	Cepheid ONewark OLo	di IVD (B2)
904 Caribbean Drive	904 Caribbean Drive	
Sunnyvale, CA 94089	Sunnyvale, CA 94089	
USA	USA	
Functional Testing		
Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed
☐ If checked this documen	at is produced electronically and valid without a wet signat	ure.
MMM	030ct22	
Signature of Quality Assu	irance, Date	
Name: Molly Doan		
Title: Quality Systems Specialist		