

Product Name: Xpert® TV

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Cepheid Catalogue Part No.: GXTV-10	▼		
Kit Lot No.: 1001463452			
Cartridge Lot No.: 22201			
Kit Expiration Date: 2026 05 31			
Legal Manufacturer Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA	Manufacturing Facility Cepheid 121 N Guild Avenue Lodi, CA 95240 USA	Solna Lodi	Sunnyvale
Functional Testing according to D206	18, Rev. M		

Test Description	Acceptance Criteria	Test Result
LOW Positive	TV detected	Passed
HIGH Positive	TV detected	Passed
Negative	TV not detected	Passed

If checked, this document is produced electronically and therefore valid without a wet signature

13 DEC 2024

Signature of Quality Assurance,

Date

Name: Nhat Dao

Title: Quality Systems Specialist

PN 301-6251 Rev. B