CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® MRSA

Cepheid Catalogue Part No.: GXMRSA-100N-10

Kit Lot No.: 1000984397

Cartridge Lot No.: 24006

Kit Expiration Date: 2025-02-23

Legal Manufacturer
Cepheid
904 Caribbean Drive
Sunnyvale, CA 94089 USA

Manufacturing Facility
Cepheid
121 N Guild Avenue
Lodi, CA 95240
USA

○ Solna ○ Sunnyvale

◉ Lodi

Functional Testing according to D16895, Rev. AA

Test Description	*Acceptance Criteria*	*Test Result*
Positive	MRSA Positive	Passed
Negative	MRSA Negative	Passed

☐ If checked, this document is produced electronically and therefore valid without a wet signature

6/14/2023

Signature of Quality Assurance, **Date**

Name: Mai Kue

Title: Quality System Specialists