

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Devices Regulations.

Product Name:

Xpert® Xpress CoV-2/Flu/RSV plus

Cepheid Catalogue Part No.:

XP3COV2/FLU/RSV-10

Kit Lot No .:

1000328101

Cartridge Lot No.:

02013

Kit Expiration Date:

2022-09-04

Legal Manufacturer

Manufacturing Facility

Solna Sunnyvale Newark

Cepheid

904 Caribbean Drive

Sunnyvale, CA 94089 USA

Cepheid

904 Caribbean Drive Sunnyvale, CA 94089

USA

Functional Testing

Test Description	Acceptance Criteria	Test Result
Negative	SARS-CoV-2 NEGATIVE;Flu A NEGATIVE;Flu B NEGATIVE;RSV NEGATIVE	Passed
Positive	SARS-CoV-2 POSITIVE;Flu A POSITIVE;Flu B POSITIVE;RSV POSITIVE	Passed

If checked this document is produced electronically and valid without a wet signature.

Name: Emaela S. Quan-Tong

Title:

Quality Assurance Supervisor