

CERTIFICATE OF ANALYSIS

This certificate is provided to confirm that the diagnostic kit specified below conforms to the production and testing specifications required by Cepheid's Quality System, in compliance with the US Food and Drug Administration's Quality System Requirements, ISO 13485, European IVD Directive and the Canadian Medical Devices Regulations (CMDR).

Product Name: Xpert® Carba-R

Catalogue Part No.: GXCARBAR-CE-10

Kit Lot No.: 1001451713

Cartridge Lot No.: 12004

Kit Expiration Date: 2026-04-05

<u>Legal Manufacturer</u> <u>Manufacturing Facility</u> • Solna Sunnyvale

Cepheid 904 Caribbean Drive Sunnyvale, CA 94089 USA

Cepheid AB Röntgenvägen 5 SE-171 54 Solna Sweden

Functional Testing

| Test Description | Acceptance Criteria | Test Result |
|------------------|---|-------------|
| LOW Positive | IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED | Passed |
| HIGH Positive | IMP1 DETECTED; VIM DETECTED, NDM DETECTED; KPC DETECTED; OXA48 DETECTED | Passed |
| Negative | IMP1 NOT DETECTED; VIM NOT DETECTED; NDM NOT DETECTED; KPC NOT DETECTED; OXA48 NOT DETECTED | Passed |

[✓] If checked this document is produced electronically and valid without a wet signature.

| Signature of Quality Assurance | Date |
|--|------|
| Alexander Avramidis (Oct 15, 2024 09:59 GMT+2) | |
| Alexander Avramidis | |

Name: Alexander Avramidis

Title: QA Analyst